

OLIGOBIND® APC blood collection tubes

Product no. ADG855T

Storage: 2 – 8°C

For Research Use Only!

PRODUCT INSERT ENGLISH

INTENDED USE

OLIGOBIND® APC blood collection tubes are intended for the collection of whole blood for further preparation and assaying of activated Protein C (APC) plasma levels. These tubes are intended for research use only.

DESCRIPTION

OLIGOBIND® APC Blood Collections Tubes are non-sterile plastic tubes that contain citrate and reversible active site inhibitors of APC. The inhibitor protects APC from inhibition by plasma protease inhibitors in drawn blood ex vivo.

APPLICATIONS

Blood samples collected using OLIGOBIND® APC Blood Collection Tubes are suitable for assaying plasma APC levels using the OLIGOBIND® APC assay (ADG855). It is not recommended to use APC tubes for other functional or antigenic coagulation assays unless the assay protocol is specifically designed to accommodate the use of these tubes.

EQUIPMENT REQUIRED BUT NOT PROVIDED

Sterile (winged) blood collection / infusion set with tubing (up to 30 cm) and short (up to 20 mm) cannula with an inner diameter of at least 0.8 mm (21G) but preferable 1.2 mm (18G).

SPECIAL PRECAUTIONS

Store OLIGOBIND® APC Blood Collection tubes at 2-8°C until use. These tubes are not sterile and should not be used for blood collection via direct venipuncture/phlebotomy. Blood should be collected using a sterile collection/infusion set with tubing (as mentioned above).

USAGE

1. Collect blood into OLIGOBIND® APC Blood Collection Tubes:
 - puncture vein by using sterile blood collection system, such as a wing canula with tubing ("butterfly") and appropriate Luer-Lock adapters
 - **note: the first blood draw (up to 3 ml) should not be used for APC determination**, therefore use other blood collection tube (up to 3 ml), e.g. for other analysis than APC determination.
 - then immediately draw blood into APC Blood Collection Tube(s).
 - discard material in case of initially unsuccessful puncture attempt or obviously impaired blood flow during sampling.
 - do not puncture the same vein again but preferably chance to other limb for second attempt.
2. Drawn blood should be stored cooled (4 °C) and centrifuged (see step 3) within 4 hours.
3. Centrifuge the blood sample at 2,500 x g for 15 minutes.
4. Plasma should be stored at 2°- 8°C and assayed within 4 hours. Alternatively, plasma may be stored at -20°C for up to 6 months.
5. Frozen plasma should be thawed rapidly at 37°C.

Follow the general guidelines for plasma preparation and handling of potentially infectious specimens.

Blood-to-additive variation can cause erroneous results if the tubes are over filled or under filled.

PRESENTATION

OLIGOBIND® APC Blood Collection Tubes are available in boxes of 25 tubes (ADG855T25) and 50 tubes (ADG855T50). The tubes have a fill size of 3 ml.

STORAGE AND STABILITY

Store OLIGOBIND® APC Blood Collection Tubes at 2°-8°C. Tubes may be used until the recorded expiration date.

BIBLIOGRAPHY

- Monitoring of plasma levels of activated protein C using a clinically applicable oligonucleotide-based enzyme capture assay. Müller J, et al., J Thromb Haemost. 2012 Mar;10(3):390-398

Distributed by:

LOXO GMBH

IMMUNOLOGIE • MOLEKULARBIOLOGIE
 BIOCHEMIE • PRODUKTE UND SYSTEME

Gerhart-Hauptmann-Str. 48
 69221 Dossenheim

Tel +49 6221 868023
 Fax +49 6221 8680255

www.loxo.de - info@loxode

Hinweis/Note:

Der Packungsbeileger dient nur als erste Information. Der relevante Packungsbeileger liegt der Ware bei.

The datasheet is for information purposes only. The current datasheet will be enclosed with product shipment.

ADG855T@01022023